


S1 (arrow marks): Tp_avg>97.5%@ 735-755 nm + Rs_avg>97.5%@ 735-755 nm, AOI=45° +/-1° S2: ARp < 0.1% @ 735-755 nm, AOI 45 deg




SIDE MEASURED: S1+S2 (good component)

COMMENTS:

OPTOMENAS UAB | Ukmerges st. 427, LT-14185, Vilnius, Lithuania | info@optoman.com | Company code: 304534353 | VAT code: LT100010934110 Luminor Bank AB, Vilnius, Lithuania | Bank Code: 40100 | IBAN: LT334010051003938875 | SWIFT: AGBLLT2X



S1 (arrow marks): Tp_avg>97.5%@ 735-755 nm + Rs_avg>97.5%@ 735-755 nm, AOI=45° +/-1° S2: ARp < 0.1% @ 735-755 nm, AOI 45 deg



SIDE MEASURED: S1+S2 (good component)

COMMENTS: Optimal coating performance is in 45-46deg range.